

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	30.3 °C	0.007 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	31.1 %HR	0.837 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	48.2 %HR	-1.859 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.9 %HR	1.452 %HR	± 3.0 %HR	± 1.9E+00 %HR