

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.862 °C	29.7 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	33.2 %HR	2.988 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.836 %HR	49.3 %HR	-1.536 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	68.7 %HR	-1.441 %HR	± 3.0 %HR	± 1.9E+00 %HR