

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.862 °C	29.8 °C	-0.062 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	30.1 %HR	-0.112 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.836 %HR	50.4 %HR	-0.436 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	70.4 %HR	0.259 %HR	± 3.0 %HR	± 1.9E+00 %HR