

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.862 °C	29.8 °C	-0.062 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	29.8 %HR	-0.412 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.836 %HR	51.6 %HR	0.764 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	69.2 %HR	-0.941 %HR	± 3.0 %HR	± 1.9E+00 %HR