

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.862 °C	30.2 °C	0.338 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	30.3 %HR	0.088 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.836 %HR	51.1 %HR	0.264 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	69.8 %HR	-0.341 %HR	± 3.0 %HR	± 1.9E+00 %HR