

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.862 °C	30.0 °C	0.138 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.212 %HR	31.6 %HR	1.388 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.836 %HR	49.9 %HR	-0.936 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	69.4 %HR	-0.741 %HR	± 3.0 %HR	± 1.9E+00 %HR