

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.293 °C	30.1 °C	-0.193 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	31.3 %HR	1.037 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.059 %HR	50.1 %HR	0.041 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.7 %HR	0.252 %HR	± 3.0 %HR	± 1.9E+00 %HR