

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.363 °C	30.4 °C	0.037 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	34.0 %HR	-1.683 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	49.0 %HR	-1.478 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.0 %HR	-0.448 %HR	± 5.0 %HR	± 1.9E+00 %HR