

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	29.3 °C	-0.993 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	25.2 %HR	-5.166 %HR	± 0.5 %HR	± 1.9E+00 %HR
50 %HR	50.366 %HR	47.1 %HR	-3.266 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.7 %HR	-1.748 %HR	± 3.0 %HR	± 1.9E+00 %HR