

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1 mm	1.01 mm	10 μm	± 40 μm	±	5.8E+01 μm
1.5 mm	1.50 mm	0 μm	± 40 μm	±	5.8E+01 μm
2 mm	2.01 mm	10 μm	± 40 μm	±	5.8E+01 μm
2.5 mm	2.50 mm	0 μm	± 40 μm	±	5.8E+01 μm
3 mm	3.01 mm	10 μm	± 40 μm	±	5.8E+01 μm
3.5 mm	3.50 mm	0 μm	± 40 μm	±	5.8E+01 μm
4 mm	4.00 mm	0 μm	± 40 μm	±	5.8E+01 μm
5 mm	5.00 mm	0 μm	± 40 μm	±	5.8E+01 μm