

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.021 °C	14.9 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.3 °C	0.051 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.369 °C	31.8 °C	-0.569 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.5 %HR	1.930 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.3 %HR	-1.148 %HR	± 3.0 %HR	± 1.9E+00 %HR