

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.780 °C	14.2 °C	-0.580 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	24.838 °C	23.6 °C	-1.238 °C	± 1.0 °C	± 4.0E-01 °C
30 °C	29.822 °C	28.1 °C	-1.722 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	35.006 °C	33.0 °C	-2.006 °C	± 1.0 °C	± 4.0E-01 °C
40 %HR	39.057 %HR	43.0 %HR	3.943 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	49.251 %HR	53.0 %HR	3.749 %HR	± 5 %HR	± 1.9E+00 %HR
60 %HR	59.517 %HR	63.0 %HR	3.483 %HR	± 5 %HR	± 1.9E+00 %HR
80 %HR	81.869 %HR	85.0 %HR	3.131 %HR	± 5 %HR	± 1.9E+00 %HR