

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.007 °C	10.2 °C	0.193 °C	± 0.4 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.0 °C	-0.021 °C	± 0.4 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.5 °C	0.251 °C	± 0.4 °C	± 4.0E-01 °C
32 °C	32.369 °C	32.1 °C	-0.269 °C	± 0.4 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.7 %HR	1.130 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	70.8 %HR	0.352 %HR	± 4.0 %HR	± 1.9E+00 %HR