

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.007 °C	10.2 °C	0.193 °C	± 0.4 °C	± 4.0E-01 °C
15 °C	15.021 °C	15.0 °C	-0.021 °C	± 0.4 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.3 °C	0.051 °C	± 0.4 °C	± 4.0E-01 °C
32 °C	32.369 °C	32.0 °C	-0.369 °C	± 0.4 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	30.7 %HR	0.130 %HR	± 4.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.2 %HR	0.752 %HR	± 4.0 %HR	± 1.9E+00 %HR