

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	30.4 °C	0.107 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	29.0 %HR	-1.693 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.570 %HR	47.0 %HR	-3.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.0 %HR	-2.448 %HR	± 5.0 %HR	± 1.9E+00 %HR