

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.293 °C                  | 30.4 °C            | 0.107 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.693 %HR                 | 29.0 %HR           | -1.693 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.570 %HR                 | 47.0 %HR           | -3.570 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.448 %HR                 | 67.0 %HR           | -3.448 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |