

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.293 °C	30.4 °C	0.107 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	29.0 %HR	-1.693 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.570 %HR	48.0 %HR	-2.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.0 %HR	-1.448 %HR	± 5.0 %HR	± 1.9E+00 %HR