

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.293 °C	29.9 °C	-0.393 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	26.0 %HR	-4.693 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.570 %HR	43.0 %HR	-7.570 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	68.0 %HR	-2.448 %HR	± 5.0 %HR	± 1.9E+00 %HR