

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.433 °C	28.7 °C	-1.733 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.693 %HR	32.0 %HR	1.307 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.713 %HR	52.0 %HR	1.287 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	75.0 %HR	4.552 %HR	± 5.0 %HR	± 1.9E+00 %HR