

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	31.807 °C	30.4 °C	-1.407 °C	±	1 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	29.0 %HR	-1.509 %HR	±	3 %HR	± 1.9E+00 %HR
50 %HR	48.024 %HR	46.2 %HR	-1.824 %HR	±	3 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	68.3 %HR	-2.015 %HR	±	3 %HR	± 1.9E+00 %HR