

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.290 °C	29.4 °C	0.110 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.265 %HR	33.2 %HR	1.935 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.153 %HR	51.5 %HR	0.347 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	68.7 %HR	-1.615 %HR	± 3.0 %HR	± 1.9E+00 %HR