

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.290 °C	29.7 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	31.265 %HR	31.9 %HR	0.635 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.153 %HR	49.8 %HR	-1.353 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	68.7 %HR	-1.615 %HR	± 3.0 %HR	± 1.9E+00 %HR