

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C             | 9.766 °C          | 9.5 °C    | -0.266 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.838 °C         | 24.8 °C   | -0.038 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.102 °C         | 30.0 °C   | -0.102 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 40.020 °C         | 39.8 °C   | -0.220 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

