

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.766 °C	9.4 °C	-0.366 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.838 °C	24.7 °C	-0.138 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.102 °C	29.7 °C	-0.402 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.020 °C	39.6 °C	-0.420 °C	± 0.5 °C	± 4.0E-01 °C

