

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.766 °C	10.2 °C	0.434 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.838 °C	25.2 °C	0.362 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.102 °C	29.9 °C	-0.202 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.020 °C	39.7 °C	-0.320 °C	± 0.5 °C	± 4.0E-01 °C

