

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)		Incertidumbre (<i>Uncertainty</i>)
30 °C	31.807 °C	30.4 °C	-1.407 °C	±	1 °C	± 4.0E-01 °C
30 %HR	30.509 %HR	26.7 %HR	-3.809 %HR	±	5 %HR	± 1.9E+00 %HR
50 %HR	48.024 %HR	52.0 %HR	3.976 %HR	±	5 %HR	± 1.9E+00 %HR
70 %HR	70.315 %HR	74.6 %HR	4.285 %HR	±	5 %HR	± 1.9E+00 %HR