

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.210 °C	-24.7 °C	0.510 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.412 °C	-0.1 °C	0.312 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	4.662 °C	4.9 °C	0.238 °C	± 0.45 °C	± #DIV/0! °C
10 °C	9.486 °C	9.7 °C	0.214 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.028 °C	24.8 °C	-0.228 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.223 °C	29.9 °C	-0.323 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.100 °C	39.7 °C	-0.400 °C	± 0.45 °C	± 4.0E-01 °C

