

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	31.368 %HR	33.0 %HR	1.632 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	52.032 %HR	53.9 %HR	1.868 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.916 %HR	67.9 %HR	-2.016 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.161 %HR	85.8 %HR	-4.361 %HR	± 5 %HR	± 1.9E+00 %HR

