

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.210 °C	-24.5 °C	0.710 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.412 °C	-0.1 °C	0.312 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	4.662 °C	5.0 °C	0.338 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	9.486 °C	9.8 °C	0.314 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.2 °C	0.172 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.223 °C	30.5 °C	0.277 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.100 °C	40.3 °C	0.200 °C	± 0.45 °C	± 4.0E-01 °C

