

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.230 °C	-24.9 °C	0.330 °C	± 0.7 °C	± 4.0E-01 °C
0 °C	-0.141 °C	-0.1 °C	0.041 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.783 °C	4.7 °C	-0.083 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.736 °C	9.6 °C	-0.136 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.938 °C	24.8 °C	-0.138 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.072 °C	29.8 °C	-0.272 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.839 °C	39.6 °C	-0.239 °C	± 0.5 °C	± 4.0E-01 °C

