

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.230 °C        | -24.9 °C  | 0.330 °C      | ± 0.7 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.141 °C         | 0.0 °C    | 0.141 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 5 °C              | 4.773 °C          | 4.8 °C    | 0.027 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 10 °C             | 9.736 °C          | 9.8 °C    | 0.064 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 25 °C             | 24.938 °C         | 25.0 °C   | 0.062 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 °C             | 30.072 °C         | 30.2 °C   | 0.128 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 40 °C             | 39.839 °C         | 39.6 °C   | -0.239 °C     | ± 0.5 °C             | ± 4.0E-01 °C                |

