

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
0 mm	0.00 mm	0 μm	± 40 μm	±	5.8E+00 μm
2 mm	2.09 mm	90 μm	± 40 μm	±	5.8E+00 μm
5 mm	4.94 mm	-60 μm	± 100 μm	±	5.8E+00 μm
10 mm	9.93 mm	-70 μm	± 200 μm	±	5.8E+00 μm
15 mm	14.99 mm	-10 μm	± 200 μm	±	5.8E+00 μm
20 mm	19.91 mm	-90 μm	± 400 μm	±	5.8E+00 μm
25 mm	25.03 mm	30 μm	± 500 μm	±	5.8E+00 μm
50 mm	50.04 mm	40 μm	± 1000 μm	±	5.8E+00 μm
100 mm	100.04 mm	40 μm	± 2000 μm	±	5.8E+00 μm