

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-5 °C	-5.806 °C	-5.5 °C	0.306 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.179 °C	25.3 °C	0.121 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	50.619 °C	51.1 °C	0.481 °C	± 1 °C	± 1.1E+00 °C
20 %HR	18.811 %HR	16.7 %HR	-2.111 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	48.944 %HR	47.1 %HR	-1.844 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	89.875 %HR	87.2 %HR	-2.675 %HR	± 3 %HR	± 1.9E+00 %HR