

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.30 μm	23.2 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
51.60 μm	51.5 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
124.60 μm	124.5 μm	-0.10 μm	± 4 μm	±	5.8E-01 μm
250.30 μm	250.2 μm	-0.10 μm	± 8 μm	±	5.8E-01 μm
486.00 μm	486.1 μm	0.10 μm	± 15 μm	±	5.8E-01 μm
963.00 μm	963.1 μm	0.10 μm	± 29 μm	±	6.0E-01 μm