

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
24.2.0 μm	23.4 μm	-0.8 μm	± 2.5 μm	±	5.8E-01 μm
50.4 μm	49.3 μm	-1.1 μm	± 2.5 μm	±	1.0E+00 μm
126.1 μm	125.1 μm	-1.1 μm	± 2.5 μm	±	1.3E-01 μm
250.6 μm	246.8 μm	-3.8 μm	± 2.5 μm	±	5.8E-01 μm
497.0 μm	492 μm	-5.0 μm	± 2.5 μm	±	4.6E+00 μm
974.0 μm	970 μm	-4.3 μm	± 2.5 μm	±	6.4E+00 μm