

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
24.2.0 μm	23.4 μm	-0.8 μm	± 2.5 μm	±	5.8E-01 μm
50.4 μm	49.3 μm	-1.1 μm	± 2.5 μm	±	1.0E+00 μm
126.1 μm	125.1 μm	-1.1 μm	± 2.5 μm	±	1.3E-01 μm
250.6 μm	247.3 μm	-3.3 μm	± 2.5 μm	±	5.8E-01 μm
497.0 μm	495 μm	-2.3 μm	± 2.5 μm	±	1.7E+00 μm
974.0 μm	973 μm	-1.3 μm	± 2.5 μm	±	6.0E-01 μm