

Certificado No. (Certificate #):		53969		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-02-13		Marca (Brand):		Insize	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		10 mm	10.01 mm	10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		20 mm	20.01 mm	10 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		50 mm	50.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		75 mm	75.02 mm	20 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		150 mm	150.02 mm	20 µm	± 30 µm	± 7.4E+00 µm	
Interior / (Interior)		50 mm	49.99 mm	-1 µm	± 30 µm	± 8.2E+00 µm	
Interior / (Interior)		100 mm	100.01 mm	10 µm	± 30 µm	± 8.2E+00 µm	
Profundidad / (Depth)		50 mm	50.01 mm	10 µm	± 30 µm	± 7.1E+00 µm	
Profundidad / (Depth)		100 mm	100.00 mm	0 µm	± 30 µm	± 7.1E+00 µm	