

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.213 °C	29.4 °C	-0.813 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.723 %HR	29 %HR	-1.723 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.263 %HR	52 %HR	1.737 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.826 %HR	75 %HR	4.174 %HR	± 5.0 %HR	± 1.9E+00 %HR