

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.112 °C	30.3 °C	0.188 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33 %HR	2.430 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	51 %HR	0.222 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68 %HR	-2.143 %HR	± 3.0 %HR	± 1.9E+00 %HR