

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.112 °C	30.4 °C	0.288 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29 %HR	-1.770 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	49 %HR	-1.578 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68 %HR	-2.343 %HR	± 3.0 %HR	± 1.9E+00 %HR