

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.112 °C	30.2 °C	0.088 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31 %HR	0.130 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	51 %HR	0.822 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68 %HR	-2.043 %HR	± 3.0 %HR	± 1.9E+00 %HR