

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
3.0000 in	3.0000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.0625 in	3.0625 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.1000 in	3.1000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.1250 in	3.1250 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.2000 in	3.2000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.2500 in	3.2500 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.3000 in	3.3000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.4000 in	3.4000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
3.5000 in	3.5000 in	0.0000 in	± 0.0002 in	±	5.9E-05 in
3.8000 in	3.8000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
4.0000 in	4.0000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in