

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.107 °C	10.2 °C	0.093 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.021 °C	15 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.119 °C	25 °C	0.281 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.118 °C	32 °C	-0.118 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.6 %HR	1.030 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.5 %HR	-0.743 %HR	± 3.0 %HR	± 1.9E+00 %HR