

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.112 °C	29.5 °C	-0.612 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32 %HR	1.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	51 %HR	0.522 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	72 %HR	1.757 %HR	± 5.0 %HR	± 1.9E+00 %HR