

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.173 °C	30.2 °C	0.027 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.640 %HR	32.5 %HR	2.860 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	52.1 %HR	1.816 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.816 %HR	72.2 %HR	1.384 %HR	± 3.0 %HR	± 1.9E+00 %HR