

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.173 °C                  | 30.1 °C            | -0.073 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 29.650 %HR                 | 30.8 %HR           | 1.150 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.284 %HR                 | 51.4 %HR           | 1.116 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.816 %HR                 | 72.8 %HR           | 1.984 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |