

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.173 °C	30.0 °C	-0.173 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.640 %HR	30.8 %HR	1.160 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	51.1 %HR	0.816 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.816 %HR	71.5 %HR	0.684 %HR	± 3.0 %HR	± 1.9E+00 %HR