

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.173 °C	30.1 °C	-0.073 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.650 %HR	32.1 %HR	2.450 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.284 %HR	51.4 %HR	1.116 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.816 %HR	71.6 %HR	0.784 %HR	± 3.0 %HR	± 1.9E+00 %HR