

| Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|----------------------|-----------------------------|------------|
| 1 mm              | 1.01 mm   | 10 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |
| 1.25 mm           | 1.26 mm   | 10 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |
| 1.50 mm           | 1.50 mm   | 0 µm          | ± 30 µm              | ±                           | 5.8E+00 µm |
| 2 mm              | 2.01 mm   | 10 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |
| 3 mm              | 3.00 mm   | 0 µm          | ± 30 µm              | ±                           | 5.8E+00 µm |
| 4 mm              | 4.01 mm   | 10 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |
| 5 mm              | 5.02 mm   | 20 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |
| 8 mm              | 8.02 mm   | 20 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |
| 10 mm             | 10.01 mm  | 10 µm         | ± 30 µm              | ±                           | 5.8E+00 µm |