

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
52.65 μm	52.70 μm	0.05 μm	± 2.5 μm	±	5.8E-01 μm
155.23 μm	155.6 μm	0.37 μm	± 2.5 μm	±	5.8E-01 μm
257.12 μm	256.6 μm	-0.52 μm	± 2.5 μm	±	5.8E-01 μm
511.72 μm	511.8 μm	0.08 μm	± 2.5 μm	±	5.8E-01 μm